

आईसीएमआर-राष्ट्रीय जनजातिस्वास्थ्यअनुसंघान संस्थान स्वास्थ्य अनुसंघान विभाग, स्वास्थ्य एवंपरिवार कल्याणमंत्रालय, भारतसरकार

> ICMR - National Institute of Research in Tribal Health Department of Health Research, Ministry of Health and Family Welfare, Government of India

No. JBP/NIRTH/Budget/Open Tender/0-2/2018-19

Dated

.02.2019

निविदा आमंत्रण सूचना NOTICE INVITING TENDER

14 FEB 2019

निदेशक, आई.सी.एम.आर - राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान संस्थान, स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य एवं परिवार कल्याण मंत्रालय, जबलपुर द्वारा दोहरी निविदा पद्धति में दिनांक 08-03-2018 दिन शुक्रवार (पूर्वान्ह 11.00 बजे) तक मोहरबंद निविदाएं निम्नलिखित उपकरणों को क्रय करने हेतु आमंत्रित की जाती हैं, जिसे उसी दिन दिनांक 08-03-2018 शुक्रवार (अपरान्ह 02.00 बजे) को आई.सी.एम.आर - रा.ज्.स्वा.अनु.सं., जबलपुर में खोला जाएगा। विस्तृत जानकारी के लिए कृपया हमारी वेबसाइट्सः http://www.nirth.res.in and www.icmr.nic.in. पर देखें।

The Director, ICMR-National Institute for Research in Tribal Health (NIRTH), Deptt. Of Health Research, Ministry of Health & F.W. Jabalpur invites Sealed Tenders in double bid system for purchased of under mentioned equipment under Madhya Pradesh Sickle Cell Anemia Control Programme Funded by Ministry of Tribal Affairs, Government of Madhya Pradesh up-to 08-03-2019 (Friday) (11.00 AM) and same will be opened on 08-03-2019 (Friday) at 02.00 P.M. at ICMR-NIRTH, Jabalpur. For information in detail, please visit our websites: http://www.nirth.res.in and www.icmr.nic.in.

S.	Name of equipment	Qty
No.		
1	Water Purification System	01
2	Variant Hemoglobin HPLC System	01
3	D10 HPLC System	01

निदेशक DIRECTOR

आईसीएमआर-राष्ट्रीय जनजातिस्वास्थ्यअनुसंधान संस्थान स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य एवंपरिवार कल्याणमंत्रालय, भारतसरकार

> ICMR - National Institute of Research in Tribal Health Department of Health Research, Ministry of Health and Family Welfare, Government of India

No. JBP/NIRTH/Budget/Open Tender/0-2/2018-19

Dated .02.2019

निविदा आमंत्रण सूचना NOTICE INVITING TENDER

निदेशक, आई.सी.एम.आर -राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान संस्थान, स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य एवं परिवार कल्याण मंत्रातय, जबलपुर द्वारा दोहरी निविदा पद्धित में दिनांक 08-03-2019 दिन शुक्रवार (पूर्वान्ह 11.00 बजे) तक मोहरबंद निविदाएं आमंत्रित की जाती हैं, ये निविदाएं उसी कार्यदिवस यानी दिनांक 08-03-2019 शुक्रवार (अपरान्ह 02.00 बजे) आई.सी.एम.आर-राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान संस्थान में निविदाकर्ताओं अथवा उनके प्राधिकृत प्रतिनिधियों, जो प्रिन्सीपल्स/मैन्युफैक्चर्स से, भारत में उनकी सहायक/सम्बद्ध कंपनियों अथवा उनके प्रत्यायित/प्राधिकृत प्रतिनिधियों के माध्यम से निविदाएं खोलने के समय उपस्थित होंगे, के समक्ष खोली जाएंगी। ये निविदाएं इस रूप में होनी चाहिए (1) तकनीकी निविदा जिसमें अनुलग्नक 'ए' में दिए अपेक्षित विनिर्देशों के अनुसार विस्तृत तकनीकी विनिर्देशों का समर्थन करने वाली विनिर्माता की मूल डेटा-शीट एक लिफाफे में एवं निविदा प्रपत्र की शर्तों के अंतर्गत चाही गई सम्बद्ध व्यावसायिक जानकारियां साथ में बयाना जमा-राशि फर्म की बेलेन्श-शीटें, आयकर चुकता प्रमाण-पत्र, उपयोगकर्ताओं की सूची आदि एक दूसरे लिफाफे में प्रस्तुत करना आवश्यक है तथा (2) वित्तीय निविदा- चालू वित्तीय-वर्ष 2018-19 के दौरान निम्नलिखित उपस्कर/सामग्री की कीमत, संस्थापन, अतिरिक्त वारण्टी/ वार्षिक अनुरक्षण अनुबंध (ए.एम.सी.) एवं प्रवर्तन के लिए।

The Director, ICMR-National Institute for Research in Tribal Health (NIRTH) (ICMR), Deptt. Of Health Research, Ministry of Health & F.W. Jabalpur invites Sealed Tenders in DOUBLE BID SYSTEM upto 08-03-2019 (Friday) (11.00 AM) and same will be opened on same working day i.e. 08-03-2019 (Friday) at 02.00 P.M. at ICMR-NIRTH, Jabalpur before the tenderers or their authorized representatives who will be present at the Principals/ opening of tenders from Manufacturers, subsidiaries/affiliates or their accredited/ authorized representatives in India. The bids should in the form of (1) TECHNICAL BID providing relevant detailed technical specifications as required at Annexure 'A' together with manufacturer's original data-sheets supporting the specifications in one cover and allied commercial details as required in terms of the tender enquiry documents alongwith Earnest Money Deposits, balance sheets, Income tax clearance certificate, users list etc. in another cover are essential and (2) FINANCIAL BID for cost of equipment, training, installation, additional warranty/AMC and commissioning etc. of the following equipments/items during the current financial year 2018-19.

S.	Name of equipment	Qty
No.		
1	Water Purification System	01
2	Variant Hemoglobin HPLC System	01
3	D10 HPLC System	01

विस्तृत निविदा प्रपत्र (अ-हस्तांतरणीय) आई.सी.एम.आर -राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान संस्थान, जबलपुर की वेबसाइट <a href="http://www.nirth.res.in"/http://www.nirth.res.in"/http://www.nirth.res.in से अथवा भारतीय आयुर्विज्ञान अनुसंधान परिषद मुख्यालय, नई दिल्ली की वेबसाइट <a href="www.icmr.nic.in"/www.icmr.nic.in"/www.icmr.nic.in से डाउनलोड किए जा सकते हैं। निविदा प्रपत्र प्रस्तुत करने के समय प्रत्येक सामग्री के लिए निविदा फॉर्म की कीमत के रूप में अलग से Director, NIRTH, JABALPUR के पक्ष में देय रु. 500/- प्रत्येक के लिए डी. डी. द्वारा भृगतान करना होंगे, अन्यथा निविदा पर विचार नहीं किया जाएगा।

Detailed tender documents can be downloaded from the website of ICMR-NIRTH www.icmr.nic.in. Payment of Rs. 500.00 as cost of tender documents for each item will have to be paid separately by D.D. in favour of Director, National Institute for Research in Tribal Health (NIRTH), Jabalpur at the time of submission of tender documents, otherwise tender will not be considered.

निदेशक DIRECTOR



आईसीएमआर—राष्ट्रीय जनजातिस्वास्थ्यअनुसंघान संस्थान स्वास्थ्य अनुसंघान विभाग, स्वास्थ्य एवंपरिवार कल्याणमंत्रालय, भारतसरकार

> ICMR - National Institute of Research in Tribal Health Department of Health Research, Ministry of Health and Family Welfare, Government of India

<u>निविदा प्रपत्र</u> Tender Document

विषय-सामग्री Contents :

क्र.सं.	विवरण	पूरी करते हैं	अभ्युक्ति
S.No.	Particulars	या नहीं Whether fulfill or not	Remarks
1.	निविदा सूचना Tender Notice		
2.	निविदाकारों को अन्देश Instruction to Bidders		
3.	सामान्य निबंधन एवं शर्तें General Terms & Conditions		
4.	स्वदेशी प्रस्तावों के लिए निबंधन एवं शर्तें		
	Terms & Conditions for Indigenous Offers		
5.	आयातित प्रस्तावों के लिए निबंधन एवं शर्तें		
	Terms & Conditions for imported Offers		
6.	आवश्यकता संबंधी विनिर्देश अनुलग्नक - 'ए'		
	Requirement Specification Annexure-'A'		
7.	सामग्रियों का देयक		
	Bill of materials		
8.	विनिर्देशों के तकनीकी अनुपालन का प्रारूप अनुलग्नक - 'बी'		
	Format of technical compliance of specifications Annexure 'B'		
9.	विचलन विवरण प्रपत्र - अनुलग्नक 'सी'		
	Deviation statement form-Annexure 'C'		
10.	कीमत की निविदा का भाग - अन्लग्नक 'डी'		
	Part of Price Bid- Annexure 'D'		
11.	Installation, Warranty & Supply of Parts/Spares - Annexure 'D(1)'		
12.	विक्रय उपरांत सेवा के समर्थन से संबंधित ब्यौरा - अनुलग्नक 'ई'		
	After Sales Service Support Details-Annexure 'E'		
13.	वारण्टी प्रमाण-पत्र - अनुलग्नक 'एफ'		
	Warranty Certificate-Annexure 'F'		
14.	निविदा स्रक्षा/बयाना जमा राशि संबंधी प्रपत्र - अनुलग्नक 'जी'		
	Bid security/Earnest Money Deposit form-Annexure 'G'		
15.	निविदाकार द्वारा दी जाने वाली जांच-सूची - अनुलग्नक 'एच'		
	Check List to be furnished by the bidder Annexure 'H'		
16.	उपयोगकर्ताओं की सूची - अनुलग्नक 'आई'		
	List of Users-Annexure 'I'		

17.	परफॉर्मेन्स बैंक ग्यारण्टी का प्रपत्र - अनुलग्नक 'जे'
	Format of Performance Bank Guarantee Annexure 'J'
18.	मैन्डेट फार्म- अनुलग्नक 'जे'
	Mandate Form- Annexure 'K'
19.	ब्रेकअप प्रस्तुत करने के लिए सामान्य मार्गदर्शी-सिद्धांत
	General Guidelines for furnishing breakup
20.	पिछले 3 वर्षों की बेलेन्श-शीट एवं आयकर चुकता प्रमाण-पत्र की
	प्रति Balance-sheet of last 3 years and income tax clearance
	certificate
21.	अतिरिक्त जानकारी, यदि कोई है
	Additional Information, if any



आईसीएमआर-राष्ट्रीय जनजातिस्वास्थ्यअनुसंघान संस्थान स्वास्थ्य अनुसंघान विभाग, स्वास्थ्य एवंपरिवार कल्याणमंत्रालय, भारतसरकार

> ICMR - National Institute of Research in Tribal Health Department of Health Research, Ministry of Health and Family Welfare, Government of India

<u>निविदा स्चना</u> TENDER NOTICE

- 1. निदेशक, आई.सी.एम.आर राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान संस्थान, जबलपुर द्वारा दोहरी बोली पद्धित में प्रिन्सीपल्स/मैन्युफैक्चरर्स से भारत में उनकी सहायक/सम्बद्ध कंपनियों अथवा उनके प्रत्यायित/प्राधिकृत प्रतिनिधियों के माध्यम से मोहरबंद निविदाएं आमंत्रित की जाती हैं। ये बोलियां इस रूप में होनी चाहिए (1) तकनीकी निविदा जिसमें अनुलग्नक 'ए' में दिए अपेक्षित विनिर्देशों के अनुसार विस्तृत तकनीकी विनिर्देशों का समर्थन करने वाली विनिर्माता की मूल डेटा-शीट एक लिफाफे में एवं निविदा प्रपन्न की शर्तों के अंतर्गत चाही गई सम्बद्ध व्यावसायिक जानकारियां साथ में बयाना जमा-राशि फर्म की बेलेन्श-शीटें, आयकर चुकता प्रमाण-पन्न, उपयोगकर्ताओं की सूची आदि एक दूसरे लिफाफे में प्रस्तुत करना आवश्यक है तथा (2) वित्तीय निविदा- चालू वित्तीय-वर्ष 2018-19 के दौरान निम्नलिखित उपस्कर/सामग्री की कीमत, संस्थापन, अतिरिक्त वारण्टी/ वार्षिक अनुरक्षण अनुबंध (ए.एम.सी.) एवं प्रवर्तन के लिए।
- 1. The Director, ICMR-National Institute for Research in Tribal Health (NIRTH), Jabalpur invites Sealed Tenders in DOUBLE BID SYSTEM from Principals/Manufacturers, through their subsidiaries/affiliates or their accredited/authorized representatives in India. The bids should in the form of (1) TECHNICAL BID providing relevant detailed technical specifications as required at Annexure 'A' together with manufacturer's original data-sheets supporting the specifications in one cover and allied commercial details as required in terms of the tender enquiry documents alongwith Earnest Money Deposits, balance sheets, Income tax clearance certificate, users list etc. in another cover are essential and (2) FINANCIAL BID for cost of equipment, training, installation, additional warranty/AMC and commissioning etc. of the following items during the current financial year 2018-2019:-
- 2. निविदा प्रपत्र की सभी सामग्रियों के ब्यौरेवार विनिर्देश, निबंधन एवं शर्तें आई.सी.एम.आर एन.आई.आर.टी.एच. जबलपुर की वेबसाइट www.nirth.res.in से अथवा भारतीय आयुर्विज्ञान अनुसंधान परिषद मुख्यालय, नई दिल्ली की वेबसाइट www.icmr.nic.in से डाउनलोड किए जा सकते हैं, जिसके लिए निविदा प्रस्तुत करने के समय प्रत्येक सामग्री के लिए अलग-अलग निविदा प्रपत्र की कीमत रु. 500/-, जो लौटाई नहीं जाएगी, का निदेशक, 'राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान संस्थान' के पक्ष में जबलपुर में देय डिमांड ड्राफ्ट प्रस्तुत करना होगा, अन्यथा निविदा स्वीकार नहीं की जाएगी। निविदा आवेदन प्रपत्र हस्तांतरणीय नहीं हैं।
- 2 The tender documents and terms and conditions of all the items listed above can be downloaded from ICMR-NIRTH, JABALPUR web site at www.icmr.nic.in for which as a document fee of Rs. 500.00 non-refundable by a Demand Draft only in favour of "Director, National Institute for Research in Tribal Health, payable at Jabalpur separately for each item should be submitted at the time of submission of tender, otherwise tender will not be considered. Tender application forms are not transferable.

3. **नोट किए जाने हेतु महत्वपूर्ण बिंदु** 3 IMPORTANT POINTS TO BE NOTED

निविदा प्रस्तुत (जमा) करने की अंतिम तारीख	08-03-2019 दिन शुक्रवार (पूर्वान्ह 11.00 बजे)
एवं समय	08-03-2019 (Friday) 11.00 AM
Last date and Time of submission of tender	
CANADA SA CALIFORNIA DE LA CALIFORNIA DE	08-03-2019 दिन शुक्रवार (अपरान्ह 02.00 बजे)
The Charles to the first terms of the control of th	08-03-2019 (Friday) 02.00 PM
वित्तीय निविदा के खोलने की तारीख व समय	14-03-2019 दिन गुरुवार को (अपरान्ह 02.00
Opening dates and time of Financial bids	बजे)
	14-03-2019 (Thursday) 02.00 PM.
निविदा प्रपत्र जारी किए जाने एवं प्रस्तुत किए जाने	निदेशक,
हेत् पता	आई.सी.एम.आर-राष्ट्रीय जनजाति स्वास्थ्य
Address for issue and submission of tender	अन्संधान संस्थान' नागपुर रोड, पो.ऑ. गढ़ा,
documents	जबलप्र - 482003 (म.प्र.)
	Director
	ICMR-National Institute for Research in
	Tribal Health (NIRTH), Ministry of Health
	& FW, Deptt. Of Health Research, GOI,
	Nagpur road, PO-Garha, Jabalpur-482 003 (MP) (India)
प्रत्येक सामग्री/उपस्कर के लिए निविदा-प्रपत्र	रु. 500/- डिमांड ड्राफ्ट/बैंकर चैक द्वारा, जो
	'निदेशक, 'राष्ट्रीय जनजाति स्वास्थ्य अनुसंधान
शुल्क, जो लौटाया नहीं जाएगा Non-refundable fee for tender document for	संस्थान' के पक्ष में जबलपुर में देय हो।
each item/Equipment.	सस्याम या या अवस्तुर व व व सा
	Rs.500.00 by Demand Draft/Banker Cheque
	in the name of Director NIRTH,
निविदा प्रपत्र इन वेबसाइट्स पर भी उपलब्ध होंगे	JABALPUR, payable at Jabalpur.
Tender documents will also be available at	
website	इाफ्ट/बैंकर चैक दवारा रु. 500/- तकनीकी बोली
Wedsite	के साथ जमा करना हैं, अन्यथा निविदा स्वीकार
	नहीं की जाएगी।
	Tender fee of Rs.500/- will only be accepted in the form of demand draft/tender fee,
	otherwise tenders forms will not be
,	accepted.
सत्यंकार की राशि (ई.एम.डी.) (बिना किसी ब्याज	निविदा प्रपत्र में अनुलग्नक 'ए' में प्रत्येक सामग्री
के) (तकनीकी बोली के साथ भेजी जाना है)	के लिए बयाना जमा राशि दर्शाई गई है, जो
Earnest Money Deposit (EMD) (Interest-free,	डिमांड ड्राफ्ट/बैंकर चैक द्वारा 'निदेशक, 'राष्ट्रीय
refundable) (to be sent with the Technical Bid)	जनजाति स्वास्थ्य अनुसंधान संस्थान', जबलपुर,
	के नाम से जबलपुर में देय हो, (तकनीकी बोली
	के साथ एक अलग लिफाफे में संलग्न हो)
	आवश्यक रूप से प्रस्तुत करें।

Amount of EMD is mentioned against the each item in Annexure 'A' of this tender document that is required to be deposited by Demand Draft/ Banker cheque in the name of **Director**, **NIRTH**, **JABALPUR**, payable at Jabalpur (along with Technical Bid in a separate envelope) positively.

नोट :

• इंटरनेट से डाउनलोड किए गए प्रत्येक निविदा प्रपत्र के साथ प्रत्येक सामग्री के लिए निर्दिष्ट निविदा-श्ल्क भेजा जाना चाहिए। ऐसा न करने पर निविदा स्वीकार नहीं की जाएगी।

Note:

• The specified tender fee should accompany with each internet-downloaded tender document for each item. Failing which tender will not be considered.

निदेशक Director, आई.सी.एम.आर.–एन.आई.आर.टी.एच. जबलपुर ICMR-NIRTH, JABALPUR

- 4 Eligibility Criteria
- 4.1 Original Equipment Manufacturers with Sales and service support facilities in India through their branch office / authorized representative in India
- Indian agent with extensive sales (Minimum 3-5 years) and service support in India preferably in MP representing foreign principals / Original Equipment manufacturers. Firms/ companies with either of the above eligibility criteria are eligible to quote for the equipment. The Indian agency / dealership/ reseller ship certificate with details of sales and service support should be enclosed with the technical bid.
- 5 Instructions to Bidders
- 5.1 Tender should be submitted in two parts viz. Part I **Technical Bid** (without Prices) and Part II **Price Bid**.
- Formats for Bids: The tenderers are requested to give detailed Quotation in the formats mentioned in the Tender Document. The bidder should furnish the following:-
- 5.2.1 Technical Bid:
- 5.2.1.1 A detailed technical proposal supported by pictures, diagrams, design as per the RFP at 'Annexure A'.
- 5.2.1.2 A Technical Compliance Statement for each parameter and every model in the format 'Annexure B' exactly in the order as mentioned in the tender document.
- 5.2.1.3 The Deviation statement 'Annexure C' if any Tender terms and conditions are deviated by the bidder. If there are no Deviations then the bidder should record 'No Deviations' in the Deviation Statement
- 5.2.1.4 Particulars of price bid must be given as per 'Annexure D' along-with annexure D(1) for installation & warranty of parts/spares.
- 5.2.1.5 After Sales support details must be in the format 'Annexure E'
- 5.2.1.6 The warranty certificate to be given by the Manufacturer / vendor should be as per 'Annexure F'
- 5.2.1.7 If the bidder is furnishing EMD through Bank guarantee then the format given in **Annexure G** is to be used.
- 5.2.1.8 Check list as per Annexure H' and enclose with the Technical Bid.
- 5.2.1.9 List of Standard Users for a period of 3 years preceding this tender (details of required users are given in 6.28 of this document) 'Annexure-I'
- 5.2.1.10 Performance Bank Guarantee as per format given as 'Annexure-J'
- 5.2.1.11 Copy of audited balance sheets to show the turn-over with Income tax clearance certificate.
- 5.2.2 Price Bid:
- 5.2.2.1 The Price Bid for each and every item should be in the format enclosed as 'Annexure D'.
- 5.2.2.2 A schedule for AMC charges for a period of maximum 05 years after warranty period must be attached along-with the price list in the enclosed **Annexure D** (1).
- 5.3 Tender Document Fee:- Tender Document Fee is also payable by companies who have downloaded the tender documents from the website www.nirth.res.in or

http://www.icmr.nic.in. They should submit the tender along with a bank draft of Rs. 500/- payable to Director, NIRTH, JABALPUR for each item separately which should enclosed with the Technical Bid. If any bidder has purchased/ paid cost only for one Tender Document and submitted their bids for more than one item, then their bid will be accepted only for one item/equipment, which will be mentioned at Sl.No.1 in the submitted tender.

- Earnest Money Deposit:- The Technical Bid (Part-1) should be accompanied by Earnest Money Deposit as indicated in our detailed technical specifications in the form of Demand Draft/Bankers Cheque, Bank Guarantee (Format enclosed Annexure G) from any Nationalized Indian Bank drawn in favour of Director, National Institute for Research in Tribal Health, Jabalpur. Tenders not accompanied with EMD in the Technical bid will be summarily rejected. The EMD will carry no interest. EMD received after due date shall not be accepted under any circumstances. EMD will be refunded in due course to the unsuccessful tenderers. The EMD is liable to be forfeited, if the bidder withdraws or amends, impairing or derogates from the tender in any respect.
- Earnest Money Deposit (EMD) of the successful tenderers will be refunded only after completing all the tendering procedures in all respect like submission of PBG and other required formalities as committed by the tenderers in their tender and/ or during negotiation/discussion etc., if tenderer fails to complete their promises/ words within the stipulated period, EMD will be liable to be forfeited in favour of Government.
- 5.5 Exemption from payment of EMD:- Firms registered with the Central Purchase Organization and NSIC who are exempted from payment of EMD are allowed exemption from payment of EMD if the product being quoted is actually manufactured by them and the product is registered with these agencies. Firms registered with these agencies selling products of other companies and not manufacturing the products being quoted by them are not allowed exemption from payment of EMD. To avail EMD exemption, the firms should submit a legible photocopy of valid Registration Certificate of the products manufactured and registered with Central Purchase Organization and NSIC in a separate envelope along with the technical bid.
- 5.6 Bid Submission:- Completed tenders signed by the bidder on each page at the bottom, super scribing the Enquiry Number and closing date should be sent to **The Director**,
 - ICMR-National Institute for Research in Tribal Health (NIRTH)
 Govt. of India, Deptt. Health Research, Ministry of Health & Family Welfare,
 Nagpur Road, PO- Garha, Jabalpur -482003 (MP) (India)

Bid and Price Bid for each item should be clearly superscribed as "Technical Bid" and "Price Bid" respectively and both these sealed covers are to be put in a bigger cover which should also be sealed and duly superscribed alongwith the Address of the office. Quotations/ Tenders delivered personally should be put in the Tender Box kept at Stores Section of ICMR-NIRTH, JABALPUR, in case of any extra ordinary tender document same may be handed over to Assistant or Section Officer of Stores Section by taking receipt. The tenderers themselves should ensure that their tenders with all requirements reach this office within the stipulated date and time. No late-tenders or delayed-tenders will be accepted. ICMR-NIRTH, JABALPUR will not be responsible for non-receipt / postal delays.

- Bid Submission last date and time: The last date and time for submission of both technical and Price bids is 08-03-2019 (Friday) at (11:00 AM). Separate envelop must be used for each equipment, incase more than one tenders are received in a single envelop, their all tenders will be summarily rejected.
- Technical Bid Opening Date and Time: The technical bid of the vendors whose tenders have been received at ICMR-NIRTH, JABALPUR before the closing date and time will be opened on same day 08-03-2019 (Friday) at (02.00 PM). Vendors or their representatives who wish to be present at the time of opening of Tech. bid may present.
- Price Bid Opening: Only those bids, which meet our technical requirement, will be considered for opening of **Price Bid**. Those vendors who qualify in the technical evaluation will be permitted to be present at the time of price bid opening. No separate communication will be made in this regard.
- Price Bid Opening Date and Time: The price bid of the vendors whose tenders have been qualified will be opened on 14.03.2019 (Thursday) at (02.00 PM) at ICMR-NIRTH, Nagpur Road, JABALPUR.
- Price Reasonability Certificate: Please quote best minimum prices. A certificate regarding the quoted rates must be given by the bidders to this effect that they have neither quoted nor sold/ supplied the same item to any other customer on lesser rates than those offered to ICMR-NIRTH, JABALPUR.
- 5.13 Language of Bid: The bid prepared by the bidder, the technical brochures and all correspondence relating to the bid shall be written in **English or Hindi** language only.
- Validity of Bid: Bids shall remain valid for 90 days after the date of opening at ICMR-NIRTH, JABALPUR. A bid valid for a shorter period may be rejected as non responsive by ICMR-NIRTH, JABALPUR. In exceptional circumstances ICMR-NIRTH, JABALPUR may solicit the vendor's consent to an extension of the period of validity. The EMD provided should also be suitably extended. The vendor accepting the request for extension will not be permitted to modify the bid.

- 5.15 INTERIM CORRESPONDENCE WILL NOT BE ENTERTAINED BEFORE PLACING OF THE ORDER.
- 5.16 Tender Documents are not transferable under any circumstances. Separate tender form is required for each item.
- 5.17 The bid should be complete in all respects otherwise liable to be rejected. Tender should be duly signed. Incomplete and unsigned bids will not be considered.
- 5.18 Quotations must be clearly written or typed without any overwriting. The person signing the bid must attest all corrections / over writings if any.
- 5.19 All rates / prices in Price bid should be given in figures and in words. If there is any discrepancy between the Prices given in figures and words the higher of the two shall be taken for our comparison.
- 5.20 Additional facilities/services, if any, offered by firm will also be taken while comparison of rates.
- 5.21 Installation of Equipments: Equipments will be installed at ICMR-NIRTH Jabalpur (M.P.).
- 5.22 **Director**, **ICMR-NIRTH**, **JABALPUR** reserves the right to defer the above dates or cancel the Tender in part or whole without assigning any reason what so ever.
- 5.23 All the terms and conditions of the tender/quotation will be dealt as per instructions/ guidelines of Central Vigilance Commission.

6 GENERAL TERMS AND CONDITIONS

- Printed terms and conditions of the firms sent along with offer, if any shall not be binding on us, if they are in full or in part are not consistent with the terms and conditions specified in this Tender.
- 6.2 Supply means Supply, installation, and commissioning and satisfactory demonstration of the whole system as per the specification ordered for.
- 6.3 Delivery:- Generally the delivery of the goods should be made within 6-8 weeks from the date of placement of purchase order. The delivery time for the equipment quoted by you should be clearly mentioned.
- 6.3.1 Liability for Late Delivery and Non Delivery:- If the Vendor fails to deliver the Goods within the time as per purchase order, a sum equivalent to 1 (one) percent of the Order Value for each week of delay or part thereof until actual delivery, will be deducted from the Contract Value, as liquidated damages, subject to maximum

deduction of 5 percent of the Order Value.

- 6.3.2 For delay in installation beyond stipulated period indicated in the supply order will attract penalty @ 0.5% per week with a maximum of 5 % (beyond mutually agreed Installation & Commissioning schedule).
- In case the supplier/agent defaults in the supply of goods as per the order placed by ICMR-National Institute for Research in Tribal Health (NIRTH), Nagpur Road, Jabalpur (MP), India fails to deliver the goods within the stipulated time as indicated in the supply orders or even within the extended delivery date, the supply order lying with the supplier/agent should stand automatically cancelled and ICMR-NIRTH, JABALPUR shall be at liberty to get the goods supplied from other available sources at the risk and cost of the defaulting supplier/agent. The additional expenditure so incurred for getting the goods shall be borne by the defaulting supplier/agent in addition to forfeiture of Earnest Money Deposit and if he fails to compensate the entire loss and if the Earnest Money Deposit are not sufficient, action will be taken to realize the loss.
- 6.5 Inspection: ICMR-NIRTH, JABALPUR may decide to inspect the goods in the following modes.
- 6.5.1 Pre shipment inspection may be done by the Personnel of ICMR-NIRTH, JABALPUR or through any reputed Inspection agency.
- 6.6 **ICMR-NIRTH, JABALPUR** shall inspect and test and where necessary reject the goods after the goods arrival at **ICMR-NIRTH, JABALPUR** in addition to Pre shipment inspection if any. The Pre shipment inspection shall in no way be limited or waived by reason of the goods having been inspected, tested and passed by **ICMR-NIRTH, JABALPUR** or the inspection agency prior to shipment.
- Packing:- The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination as indicated in the Contract. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit. Suppliers will have to make good the shortages, if any, which are revealed after opening the packages. The stores which are spoiled / damaged during transit due to faulty packing will have to be replaced by the suppliers.
- 6.8 The **Director**, **ICMR-NIRTH**, **JABALPUR** reserves the right to accept the delivery in part or in full. Damaged and unaccepted supply will be returned to the vendor at vendor's risk and cost.
- 6.9 Force Majeure clause:- The vendor shall not be liable for forfeiture of its bid security, performance security, liquidated damages or termination for default, if

and to the extent that, its delay in performance or other failure to perform its obligations under the contract is the result of an event of Force Majeure. Force Majeure means an event beyond the control of the supplier and not involving the vendor's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the vendor either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes. If a Force Majeure situation arises the Vendor shall forthwith notify ICMR-NIRTH, JABALPUR in writing of such conditions and cause thereof. Unless otherwise directed by ICMR-NIRTH, JABALPUR in writing, the vendor shall continue to perform its obligations under the contract as far as is reasonably practical, and shall seek all alternative means for performance not prevented by the Force Majeure event.

- Dimensions and weight:- The dimensions of the equipment, the net weight and gross weight of the equipment should be clearly mentioned in the technical bid.
- 6.11 Applicable Law: This tender documents and contract is subject to the Indian law and Indian Contracts Act as amended from time to time.
- 6.12 Jurisdiction:- All disputes are subject to the courts within **JABALPUR**, India jurisdiction only.
- 6.13 Full and complete set of documentation manuals, Operation manuals and service / Maintenance manuals to be provided in English.
- Resolution of Disputes:- The ICMR-NIRTH, JABALPUR and the Firm shall 6.14 make every effort to resolve amicably by direct informal negotiation any disagreement or dispute arising between them under or in connection with the Purchase. If, after thirty days from the commencement of such informal negotiations. ICMR-NIRTH, JABALPUR and the Firm have been unable to resolve amicably a Contract dispute, either party may require that the dispute be referred for resolution to the formal mechanisms as under. In the case of a dispute or difference arising between ICMR-NIRTH, JABALPUR and the Firm relating to any matter arising out of or connected with the agreement, such dispute or difference shall be referred to the award of two Arbitrators, one Arbitrator to be nominated by ICMR-NIRTH, JABALPUR and the other to be nominated by the Firm or in the case of the said Arbitrators not agreeing, then to the award of an Umpire to be appointed by the Arbitrators in writing before proceeding with the reference, and in case the Arbitrators cannot agree to the Umpire, he may be nominated by the Director General, Indian Council of Medical Research, New Delhi. The award of the Arbitrators, and in the event of their not agreeing, of the Umpire appointed shall be final and binding on the parties. The Indian Arbitration & Conciliation Act 1996, the rules there under and any statutory modification or re-enactments thereof, shall apply to the arbitration proceedings. The venue of arbitration shall be ICMR-NIRTH, JABALPUR.
- 6.15 Language of contract:- The contract / Purchase Order and all correspondence

relating to the contract / Purchase order shall be written in English / Hindi language only.

Notices and communications:- For all purposes of all notices, the following address shall be used by the Bidder / Vendor:-

The Director,

ICMR-National Institute for Research in Tribal Health (NIRTH),

Govt. of India, Deptt. of Health Research,

Ministry of Health & Family Welfare

Nagpur Road, P.O. Garha, Jabalpur - 482003 (MP), India

- 6.17 Termination for default:- **ICMR-NIRTH**, JABALPUR may, without prejudice to any other remedy for breach of contract, by written notice of default to the Vendor, terminate the contract in whole or in part:
- 6.17.1 If the vendor fails to delivery any or all the goods mentioned in the purchase order or within any extension thereof granted by **ICMR-NIRTH**, JABALPUR.
- 6.17.2 If the vendor fails to perform any other obligation under the order.
- 6.17.3 If the vendor in the judgment of ICMR-NIRTH, JABALPUR has engaged in corrupt of fraudulent practices in competing for or in executing the order.
- 6.17.4 If the vendor is declared insolvent / bankrupt in a court of law or such proceedings are contemplated by **ICMR-NIRTH**, JABALPUR.
- 6.17.5 If the vendor goes into dissolution or liquidation or transfers substantial part of its business or assets.
- 6.18 Termination for Cause:- **ICMR-NIRTH**, JABALPUR may at any time terminate the purchase order in whole or in part by giving written notice to the vendor, without compensation to the vendor.
- Intellectual Property Rights:- The Vendor shall indemnify **ICMR-NIRTH**, JABALPUR against all third party claims of infringement of Patent, Trademark, Copy Rights or Industrial Design rights arising from the use of the equipment / goods or any part thereof in India.
- 6.20 Training clause:- Specific mention should be made about the Training facility available in India or abroad. If extra, its cost should be mentioned.
- Pre installation requirements:- The Technical Bid must contain details with specification of items required for installation (PRE-INSTALLATION REQUIREMENTS (AIR / WATER) ELECTRICAL POWER / TEMPERATURE & HUMIDITY CONTROL / FOUNDATION / SPACE and any other requirements to complete Installation at ICMR-NIRTH, JABALPUR.
- 6.22 Order acknowledgment:- Within 15 days from the date of Purchase order /

Contract the Order / Contract should be acknowledged by the Vendor (Indian OEM / Foreign Principal) failing which it will be deemed that the vendor is not interested in the contract / purchase order and the bid security will be forfeited without any further recourse.

- Performance Bank Guarantee/Fixed Deposit Receipt/ Demand Draft:- In order to ensure due performance of the equipment/item costing more than 1.00 lakh, each and every successful vendor/bidder irrespective of its registration status etc. has to furnish a performance bank guarantee in the form of Account payee bank guarantee / Fixed Deposit Receipt/ Demand Draft in favour of Director, NIRTH, Jabalpur, from a nationalized bank in India, in case FDR, that should be pleged in favour of Director, NIRTH, Jabalpur also. The Bank guarantee/ Fixed Deposit Receipt/ Demand Draft should cover the warranty period plus sixty days from the date of installation of the equipment. The value of the Bank guarantee is 10% of the LC value. This PBG must be submitted prior to opening of F.L.C. For Indigenous Equipment PBG will be 10% of the ordered value of the equipments which is to be submitted before delivery of equipments.
- 6.24 The system (equipment) should be able to operate on power supply of 230V ($\pm 10\%$), 50 Hz single phase AC power supply until asked specially otherwise for a particular equipment.
- Risk Purchase Clause:- If the Supplier fails to deliver the goods within the maximum delivery period specified in the Purchase Order, JCMR-NIRTH, JABALPUR may procure, upon such terms and in such a manner as it deems appropriate, Goods or Services similar to those undelivered and the Supplier be liable to ICMR-NIRTH, JABALPUR for any excess costs incurred for such similar Goods or Services.
- 6.26 Responsibility:- The responsibility lies with the successful bidder, if any damage or loss to the property of the Institute occurs while executing the order.
- 6.27(A) Guarantee/Warranty, Defects & Liability Period:- (Warranty) The equipment should be on 3 years Comprehensive warranty and 2 years Non-comprehensive services for equipment free of cost from the date of delivery and completion of satisfactory installation and demonstration as certified by ICMR-NIRTH, JABALPUR, if cost of equipment is more than 10.00 lakhs otherwise 3 years comprehensive warranty only. The supplier should rectify free of cost any defect that may develop in the system during the warranty period. The warranty period shall automatically stand extended by the time taken to rectify the defects. The Firm should give a warranty certificate as per the format attached in the Tender Document.
- 6.27 (B) Bidder/firm should provide rates list of spare parts to maintain the equipment upto 05 years from 3rd or 5th year onwards as the case may be. Discount on all the parts should be given atleast upto 05 years, if cost of equipment is more than 10.00 lakhs. During AMC of equipment, there should be atleast 2 preventive and

unlimited breakdown visit/services in a year.

- User List:- A signed list of similar equipment / stores supplied by the bidder (or the OEM) to the Other ICMR Institutes/Centres, AIIMS, CSIR labs, ICAR, R & D Institutions like IITs, IISc, Railways, MES, State and Central Govt. Departments, Undertakings in India and abroad for the past three years, should be furnished, indicating capacity, model, type of equipment, cost etc. and address of the parties to whom supplied.
- The acceptance of the tender/quotation will rest with the **Director**, **ICMR-NIRTH**, **JABALPUR** who does not bind him/herself to accept the lowest quotation and reserves the right to him/herself to reject, or partially accept any or all the quotations received, without assigning any reason. The Director also reserves the right to modify/relax any or more terms & conditions of the tender. The quotations are liable to be rejected, if any of the conditions are not complied with.
- 6.29.2 The **Director**, **ICMR-NIRTH** also reserves the right to award the works/supply order or not on quality/technical basis, which depends on quality/capability of the system and infrastructure of the firm. The **ICMR-NIRTH** will not be under obligation to give clarifications in this regard and decision of the **Director** taken on this regard will be final.
- 6.30 The bidder shall supply the following documents at the time of delivery/installation:
- 6.30.1 Complete set of Technical/Operation and Maintenance Manual.
- 6.30.2 An inventory of items delivered.
- 6.30.3 Comprehensive inventory.
- 6.30.4 Node & connectivity details.
- 6.30.5 Access Points, Controller and other equipments installation Report.
- 6.30.6 System diagram.
- 6.30.7 Certification details of Access Points, Controller and other equipments.
- 6.30.8 Firm which will not fulfill the conditions accepted by them in their tender may be declared "black listed" and in that case all type of future business with that firm may also be stopped.
- 6.30.9 The firm, which have done turnover of minimum Rs. **Fifty lakhs** in any year during last three years will be eligible for submitting the tender form. Please submit the audited balance sheets of last three years as a proof, otherwise tender

will not be considered.

- 6.30.9.1 Prior turnover and prior experience to Micro & Small Enterprises (MSEs) is relaxable subject to meeting of quality and technical specifications in accordance with the relevant provisions of GFR-2017.
- 6.30.10 If single response is received, then inquiry may be refloated, however, second time if again single response is received, it may be considered with the permission of the Director.
- 6.30.11 Numbering in the pattern of 1 of 2, 2 of 2 and so on, should be mentioned on each and every page of tender documents submitted by the bidder and list of enclosures should also be mentioned at first page of tender where details for tender is written alongwith the number of pages of enclosures to avoid any unhappening/confusion etc.
- 6.30.12 If quoted equipment or model is discontinued by the firm/mfg. after quoting and accepting the item, then the firm will inform to the office and after getting acceptance from ICMR-NIRTH, the firm will have to supply the equivalent or higher version of the same make at the same price, other terms and conditions of the tender will remain un-changed.
- 6.30.13 All other clauses not specifically mentioned in this tender but are part of the terms and conditions of this tender shall be ipso facto applicable to this tender.
- 6.30.14 Please check regularly our (ICMR-NIRTH) website for any changes/modifications / amendments in the tender, as any changes will be uploaded only on website.

Clarification of Tenders

The ICMR-NIRTH may ask at its discretion any of the bidder any clarification of tender condition including clarification of methodology of execution of items, the rate analysis for the items and all other details as considered essential by ICMR-NIRTH. The above clarification or submission of all the details shall form part of the tender and shall be binding on the bidders.

Canvassing

The canvassing in any form in respect to this tender is strictly prohibited and will result in summarily rejection of tender of such bidders.

7 TERMS & CONDITIONS FOR INDIGENOUS OFFERS

7.1 Prices:- Prices in the Price Bid should be quoted Either Ex-works or on Free Delivery at NIRTH, JABALPUR basis. If it is Ex-works then the packing and forwarding charges, Insurance (from warehouse to warehouse with all risks covered on 110% of the value) charges, Freight charges by road (any reputed transporter) / Courier (any reputed courier) on door delivery basis must be clearly mentioned. Quotations without these details will be summarily rejected.

- 7.2 Excise duty:- This institute is exempted from payment of Excise duty as per notification No. 10/97 Central Excise, dated 1st March 1997. Hence excise duty exemption certificate will be forwarded along with the purchase order for excise free shipment.
- 7.3 High Sea Sale Agreement:- High Sea Sale Agreement will not be applicable for purchase of equipment mentioned in this tender.
- 7.4 Taxes:- In the case of indigenous item, the TAX Rate should be clearly indicated wherever, chargeable.
- 7.5 Packing and Forwarding:- The packing and forwarding charges should either be included in the Unit rate or quoted as a percentage of the PO or quoted as a lump sum amount.
- 7.6 Insurance:- If the contract requires the Vendor to insure the goods, the vendor unless otherwise agreed on the face hereof, insure the goods.
- 7.6.1 For a value equal to one hundred and ten percent (110%) of the order value on warehouse to warehouse basis.
- 7.6.2 On the basis of All Risks (Institute Cargo clauses) or on equivalent terms.
- 7.6.3 With underwriter (s) or insurance company(ies) of good repute.
- 7.6.4 Any additional insurance coverage shall be arranged by vendor upon the special request and on the account of ICMR-NIRTH, JABALPUR.
- 7.7 Provide all assistance to **ICMR-NIRTH**, **JABALPUR** in case of any claim is made on the insurance company on the basis of the policy
- 7.8 Freight:- If the consignment has to be dispatched by Road transport / courier then only reputed transporters / reputed courier are to be used by the vendor and consignment to be dispatched on Freight pre paid basis and door delivery basis.
- 7.9 Banker's address:- Our banker is **State Bank of India, Medical College Branch**, **Jabalpur** (IFSC code SBIN0001445) All cheques / demand drafts will be crossed account payee only.
- 7.10 Registration Numbers:- The following certified copies should be attached with the technical bid.
- 7.10.1 Goods & Service tax registration number (GSTN)
- 7.10.2 Current Income tax clearance & PAN
- 7.11 Service support:- The post installation warranty support is required to be provided by your *Jabalpur* / nearest place service centre. Hence mention the contact address,

with name of contact person and telephone numbers.

- Payment terms:- No advance payment shall be made. 100% Payment will be released within 20 days after successful installation of the equipment at ICMR-NIRTH, JABALPUR and on submission of performance Bank Guarantee of 10% value of the order price, valid till the expiry of the comprehensive warranty period and additional sixty days thereafter by the bidder. However, part payment may be released after submitting the proof of dispatches of goods from its premises in special circumstances. Decision of the Director will be final.
- 7.12.1 Price reasonableness will be decided on the basis of offered cost, prevailing market rates, last purchase prices, if any, economic indices of the raw material/labour, services, other input costs and intrinsic value etc.
- 7.12.2 As per GOI instructions, all payments to suppliers etc. are to be made by direct credit to the bank account of the beneficiaries through PFMS. Therefore all the beneficiaries (Suppliers/firms) are essentially required to submit/fill up the mandate form duly attested alongwith a cancelled cheque of the firm. In case there is a change in the beneficiary's account number etc. in later stage, the revised mandate form shall have to be furnished.
- 7.13 Documents:- The following documents are required. The invoice will not be paid until supported by all the supporting vouchers.
- 7.13.1 Invoice in triplicate showing the Purchase order / contract number, goods description, quantity, unit price and total amount. The invoice should be duly Pre receipted on a revenue stamp (revenue stamp mandatory vide Indian Stamp Act for amount exceeding Rs. 5000/-).
- 7.13.2 Acknowledgement of the receipt of goods from the consignee by the transporter / courier.
- 7.13.3 Insurance certificate / Premium Paid Voucher / Policy if any.
- 7.13.4 Manufacturer's warranty certificate.
- 7.13.5 Inspection certificate issued by the nominated inspection agency, if any (nominated as per contract) or the Vendor's inspection / testing report.
- 7.13.6 Packing list (two copies) identifying the contents of each package.
- 7.13.7 Delivery challan / Delivery note in triplicate

8 TERMS & CONDITIONS FOR IMPORTED OFFERS

8.1 Cost:- Ex-Works Price to be clearly indicated in the Price Bid. The Indian agents commission if any will be subtracted from the Ex-works Price in the Purchase

Order. **FOB** Charges up to the nearest gateway International airport to be indicated in the Price bid.

- High Sea Sale Agreement:- High Sea Sale Agreement will not be applicable for purchase of equipment mentioned in this tender.
- Import License: The ordered materials are being imported under Open General License. However this office has its own **IECode No.1109004486** issued from Director General of Foreign Trade, Bhopal M.P. vide file No. 11/04/130/00466/AM10 dated 18.11.2009.
- Export License:- If the ordered equipment are subject to Vendor procuring an export license from the designated government agency / Country from where the goods are shipped / sold, the vendor has to mention the name, address of the government agency / authority. The vendor must also mention the time period within which the license will be granted in normal course.
- 8.5 Customs duty: This institute is exempted from payment of Customs duty as per notification No.51/96 Customs dated 23.7.1996 and Central Excise Duty Exemption vide notification No. 10/97- Central Excise dated 01.3.1997 issued by DSIR vide letter No. TU/V/RG-CDE(243)/2016 dated 23.9.2016.
- INDIAN AGENT'S COMMISSION, if any: If you have any Agents in India, please show specifically whether the amount of commission payable to them is included in the Ex-works Price. Please note that under the Government of India regulations, Commission to Indian Agents is payable only in Indian Currency. The role played by the Indian Agents in rendering assistance to ICMR-NIRTH, JABALPUR must also be specified and copy of authorization letter enclosed in technical bid.
- 8.7 COUNTRY OF ORIGIN: Please specify PLACE / Country of Origin and the Place / country from which goods will be finally shipped. These particulars are very important for establishment of L/C, arranging insurance etc.
- Forwarding Agent:- ICMR-NIRTH, JABALPUR has an existing contract with M/S Airport Handling Services, D-14, Sector 6, Noida-201301 UP (New Delhi) Telephone No. 0120-4738165-68, 0120-4738100 for providing the services of a Freight Forwarder. All FCA / FOB shipments will be consolidated by the agents of M/S Airport Handling Services, Noida.
- 8.9 Clearing Agent:- ICMR-NIRTH, JABALPUR has an existing contract for clearance of imported cargo with M/S Airport Handling Services, D-14, Sector 6, (Noida-201301) UP. All the pre-shipment alerts and post shipment documents should also be forwarded to them.
- 8.10 Inspection:- The Indian agent / Indian branch Office will be intimated immediately after the receipt of the consignment at ICMR-NIRTH, JABALPUR Stores. The

consignment will be jointly inspected by the Stores personnel, Technical Personnel of ICMR-NIRTH, JABALPUR and the representative of Indian agent within 24 hours after receipt. The consignment will be compared with the purchase order and packing list. Any excess supply, shortage, wrong supply and any damage will be noted immediately and intimated to the vendor. Damaged and unaccepted supply will be returned to you at your risk and cost. The vendor should ship the shortage and replace the damaged / wrong shipment at no cost to ICMR-NIRTH, JABALPUR/MRHRU, Datia (Madhya Pradesh).

- 8.11 Installation and commissioning: ICMR-NIRTH, JABALPUR requires the assistance of the Indian agent/Indian branch Office during installation, commissioning of the Equipment. Training for the personnel at ICMR-NIRTH, JABALPUR/MRHRU, Datia (Madhya Pradesh) is also mandatory when specifically asked for in the contract.
- 8.12 Service: The Indian agent / Indian branch office of the foreign principal should support the equipment fully during the warranty period. If **ICMR-NIRTH**, JABALPUR requires AMC during the post warranty period the Indian agent / Indian branch office must fully offer full support.
- 8.13 Merger and Acquisition:- Any merger and acquisition of the Foreign principal by other companies / firms must be intimated to ICMR-NIRTH, JABALPUR immediately.
- 8.14 Change of Indian Agent:- If there is any change of Indian agent during the warranty period / post warranty period the foreign principal must intimate ICMR-NIRTH, JABALPUR immediately with the new company's name and address.
- 8.15 Insurance:- If the contract requires the foreign principal to insure the goods, the vendor unless otherwise agreed on the face hereof, insure the goods
- 8.15.1 For a value equal to one hundred and ten percent (110%) of the FOB amount of the contract on warehouse to warehouse basis.
- 8.15.2 On the basis of All Risks (Institute Cargo clauses) or on equivalent terms
- 8.15.3 With underwriter (s) or insurance company(ies) of good repute
- 8.15.4 Any additional insurance coverage shall be arranged by vendor upon the special request and on the account of ICMR-NIRTH, JABALPUR
- 8.15.5 Provide all assistance to **ICMR-NIRTH**, JABALPUR in case of any claim is made on the insurance company on the basis of the policy
- 8.16 Import Payment:- Letter of Credit will be established for 100%. Payment will be released in two parts:

- 8.16.1 **80**% payment after presentation of original documents after dispatch of the item ordered.
- 8.16.2 20% payment after satisfactory supply, installation and commissioning and demonstration of the item ordered and furnishing 10% Performance Bank Guarantee on any Nationalized Bank in India by Foreign Principal / Indian Agent valid till the expiry of the comprehensive warranty period and additional sixty days thereafter and all other promises committed by the tenderer in their tender and/ or during the negotiation/discussion.
- 8.16.3 If the Letter of Credit is required to be extended / reinstated for reasons not attributable to **ICMR-NIRTH**, **JABALPUR**, the Bank Charges thereof shall be to the beneficiary's account.
- 8.16.4 Documents: The following documents are required by us at the time of negotiation of documents for Letter of Credit within 48 hours of shipment.
- 8.16.4.1 Commercial invoice 3 copies (original)
- 8.16.4.2 Packing list 2 copies
- 8.16.4.3 Country of Origin 1 copy
- 8.16.4.4 Master Airway Bill 2 copies
- 8.15.4.5 Warranty certificate 2 copies

निदेशक Director, आई.सी.एम.आर.–एन.आई.आर.टी.एच. जबलपुर ICMR-NIRTH, JABALPUR

Requirement Specifications

SI. No.	Name of the Equipment	Qty Requir ed	Specification	EMD Amount (Rs.)
1	Water Purification system	01	water Purification system should be capable of independently dispensing both RO and ultrapure water. RO able to use tap water as feed (able to accept upto 2000 micro Siemens conductivity). The generated water should be treated with UV to destroy bacteria, before it is stored in the storage tank. The storage tank should be capable of maintaining the purity of stored water. Biofilm formation should be prevented Production rate minimum 20L/h. Continuous monitoring and display of water quality Parameter. Should have customizable alert and alarms. Instrument should work at ambient temperature upto 45°C. Certificate of calibration from recognized international agencies. Should meet internationally recognized safety norms. Instruments should be supplied with the inlet water tank (100 L) along with stand. Water supply pressure: 2 m head. Ultrapure water resistivity:- Pyrogens (endotoxins) <0.001EU/mL RNases<1pg/mL Resistivity (MΩ-cm) = 18.2MΩ•cm @ 25 deg C Conductivity (μS/cm) < 0.055 μS/cm TOC ≤5ppb Bacteria<10CFU/100mL DNases<5 pg/mL Particulates(size>0.22μm) <1/mL Sodium (ppb)<1 Chloride (ppb) < 1 Total Silica (ppb) < 3 Flow rate >1L/min RO water with the following property Resistivity (MΩ-cm) = 12-15 MΩ•cm @ 25 deg C, or more Conductivity (μS/cm) <1 TOC <30ppb Bacteria<10CFU/100mL Pyrogens (endotoxins) <0.001EU/mL RNases<1pg/mL DNases<5 pg/mL Particulates(size>0.22μm) <1/mL Sodium (ppb)<5, Chloride (ppb) < 5 Total Silica (ppb) < 3. Flow rate >1L/min Row rate >1L/min	50,000/-
2	Variant Hemoglobin- HPLC system	01	Input power voltage: 230 V, 50-52 Hz AC Automated HPLC system, must be dedicated to Thalassaemia and Haemoglobinopathy testing and screening. The system should be able to screen and quantitate haemoglobins Hb A2, Hb A and Hb F and detect the most commonly occurring abnormal hemoglobins like Hb S, Hb D, Hb E, Hb C, Hb Q- India, Hb D-Iran, Hb Lepore, Hb Saurashtra and other rare abnormal hemoglobins in both homozygous and in single and double heterozygous conditions. The system should have the provision of presumptive identification of Hb Barts and Hb H and various alpha chain variants like Hb J Meerut, Constant Spring etc. The company must have an installation base of minimum 200 in India for doing Thalassaemia and Haemoglobinopathy of which there should be atleast 20 installation in NABL or CAP accreditated Laboratories and should be able to provide the relevant product and service support. The company must have at least 20 years of presence in India with availability of system & reagents for thalassaemia and haemoglobinopathies testing. And	1,20,000/-

	must have users for Haemoglobinopathies for a minimum of last 15 years.
6	A minimum of 10 customer satisfaction certificate should be provided.
-	The HPLC system should have a dual piston pump so that each elution buffe.
7	has a different pump and the buffers work efficiently to give a continuous and a
,	precise buffer gradient.
	The system should not take more than 6.5 mins for the screening of thalassemia
8	
	and hemoglobinopathies. The system should have spinning of vacutainer before aspiration to avoid
9	
	improper sampling.
10	The system should have automatic barcode positioning and reading facility.
11	The system should have continuous or batch wise sample analysis with randon
11	access and sample bar code sensor
10	The system should have the facility of primary tube sampling and direct dilution
12	of the samples without manual intervention.
	The system should have Tube Venting Capabilities so that there is no resistance
13	caused while pulling blood from the tube which can impact the repeatability o
13	results.
	Complete ready to use reagent kit must be provided with buffers in plastic tank
	to view the levels of buffers during the run. Columns, primers, calibrators and
14	to view the levels of ouriers during the full. Columns, princes, canorators and
	sample vials must be within the kit as a single kit thus making it easy to
	calculate cost per test.
15	All reagents required should be of the same lot for reliability of result and cos
15	calculation per test.
	The system should have an inbuilt system check facility which checks that all
16	the system parameters (eg, cartridge, buffer, reagent, waste etc.) are read
	before the sample analysis.
	The buffers must be provided with in plastic tanks to view the levels of buffer
	during the run. Also there should be a system which monitors liquid volume b
17	weight and an alarm is generated by software if the buffer reduces than the so
	volume.
-	The system should have automated sampler module which can accommodat
18	10 sample racks together. Each sample rack is barcoded and have 10 sample
10	positions. The system should have continuous loading facility during the run.
	The system should have easy maintenance which should not incur additional
19	
	cost of purchasing cleaning material or solutions.
20	The system should not require adjustment of flow rates for maintaining of
	Retention time by the user.
	The system should have dedicated computer and software, which enables th
21	system for bidirectional interfacing. Moreover the software should have
21	customized reporting format, giving info on the subtype and quantity of
	hemoglobin detected.
22	It should have a sufficient data hard disk approx.80GB hard and a remote data
22	access feature when connected to LAN or Intranet.
	The system must have a software for real time viewing of the analysis of the
23	sample.
	It should have an offline CD-ROM and an online chromatogram library which
	should be a searchable database with more than 400 chromatograms of full
24	classified abnormal haemoglobins and thalassaemias along with their clinical
	and molecular classification.
	The HPLC system must be used in govt. thalassaemia screening programs i
1212	The HPLC system must be used in govt, thatassachina screening programs in
25	India and a minimum of 15 Govt user list of the thalassaemia kit should be
	provided.
	The system should have an on board QC Menu capable of storing the qualit
26	control data and printing the standard deviation, Coefficient of Variation value
	and LJ chart.
	The company should provide normal and abnormal third party controls for H
27	A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) t
15 55 050	help compare results with similar users worldwide.
	The system should have dedicated computer, laser printer and software, which
	enables the system for bidirectional interfacing. Moreover the software should
20	give information on the subtype and quantity of hemoglobin detected. Also the
28	software should enable result storage of minimum 10000 chromatogram
	I SOITWALC SHOULD CHADIC ICSUIT STOTAGE OF HIMBHIUM TOOOD CHIOMATOGRAM
29	(without any additional purchase of software). The result from the machine should be presented in a symmetrical order.

		(vertical) with proper description of date, time of injection, sample ID, age, sex, total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result.	
	30	It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility.	
	31	The system should be capable of holding 10 racks at a time so that it can be used for at least 100 vials at a time.	
	32	The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment.	
	33	The waste tank should be sufficiently big (atleast 5 liters) so that it reduces user interference with the machine and help in smooth running of large volume of samples without interruption.	
	34	The reagent containers should have a capacity of more than 1.5 liters so that the user does not need to change buffers regularly.	
	35	The company must have feature of capillary collection kit for remote sample	
		collection with sample stability at 2-8 C for 14 days The company must provide a support of factory trained engineers, application	4
	36	specialist and thalassemia expert for the technical and chromatogram interpretation related issues.	
	37	The HPLC system should have flexibility to use various samples tube sizes of 13x75mm, 13x100 mm micro capillary tubes (Micro cuvette startstedt), Micro capillary tubes brands (other than Micro cuvette startstedt) defined as sample	
		vials 1.5ml sample vials. The instrument provider should also provide training to the pathologist for	
	38	working, quality control, testing, evaluation and interpretation.	
3 D10-HPLC 01 system	1	Automated HPLC system, must be dedicated to Thalassaemia and Haemoglobinopathy testing and Screening and A1c testing.	41,000/-
	. 2	The system should be able to screen and quantitate haemoglobins Hb A2, Hb A, HbA1c and Hb F and detect the most commonly occurring abnormal hemoglobins like Hb S, Hb D, Hb E, Hb C, Hb Q- India, Hb D-Iran, Hb Lepore, Hb Saurashtra and other rare abnormal hemoglobins in both homozygous and in single and double heterozygous conditions.	
	3	The system should have the provision of presumptive identification of Hb Barts and Hb H and various alpha chain variants like Hb J Meerut, Constant Spring etc.	
	4	The company must have an installation base of minimum 1000 in India for doing Thalassaemia and Haemoglobinopathy of which there should be atleast 30 installation in NABL or CAP accreditated Laboratories and should be able to provide the relevant product and service support	-
	5	The company must have at least 15 years of presence in India with availability of system & reagents for thalassaemia and haemoglobinopathies testing. And must have users for Haemoglobinopathies for a minimum of last 10 years.	
	6	A minimum of 10 customer satisfaction certificate should be provided. The system should have dual performance of Hemoglobin Testing for	
	'	thalassemia and hemoglobinopathies and A1c estimation using a single kit so as	
		to avoid any change of kit or reagents for each test, thus reducing the loss of buffers and reagents. The change from one test to another should just happen with a click of button and should estimate A1c in not more than 3.5 mins and A2/F estimation in not more than 6.5 mins.	
	8	The system should have automatic barcode positioning facility. The system should have continuous or batch wise sample analysis with random	
		access and sample bar code sensor	
	10	The system should have the facility of primary tube sampling and direct dilution of the samples without manual intervention.	
	11	The system should have Tube Venting Capabilities so that there is no resistance caused while pulling blood from the tube which can impact the repeatability of results.	
	12	Complete ready to use reagent kit must be provided with buffers in plastic tanks to view the levels of buffers. Columns, primers, calibrators and sample vials must be within the kit as a single kit thus making it easy to calculate cost per	

	test.
13	All reagents required should be of the same lot for reliability of result and cost calculation per test.
14	The system should have an inbuilt system check facility which checks that all
	the system parameters (eg, cartridge, buffer, reagent, waste etc) are ready before
	the sample analysis.
1.5	The buffers must be provided with in plastic tanks to view the levels of buffers
15	during the run. Also there should be a system which monitors liquid volume
	during the run. Also there should be a system which mointors riquid volume
	and an alarm is generated by Software if the buffer reduces than the set volume
16	The system should be able to detect correct A1c values in presence of abnormal
	hemoglobin variants like HbD, HbE, HbS & HbC
17	The System should be NGSP (National Glycohemoglobin Standardisation
	Program) Certified and traceable to IFCC reference method.
18	The system should offer both NGSP & IFCC value reporting on the same
10	patient report, control & calibrator report.
19	The required item must be USFDA or European CE approved product
19	(authentic documents must be submitted along with the offer).
	(authentic documents must be submitted along with the orier).
20	The HPLC system should have better precision, CV less than 2.5%.
21	The system should have in-kit external standards for instrument calibration
	ensuring accurate quantitation of results.
22	It should have an sufficient data hard disk approx. 80GB hard drive and a
	remote data access feature when connected to LAN or Intranet.
23	The system must have a software for real time viewing of the analysis of the
-	sample
24	It should have an offline CD-ROM and an online chromatogram library which
24	should be a searchable database with more than 400 chromatograms of fully
	should be a scalehable database with more than 400 enrollate grains of tarry
	classified abnormal haemoglobins and thalassaemias along with their clinical
	and molecular classification.
25	The HPLC system must be used in govt. thalassaemia screening programs in
	India and a minimum of 15 Govt user list of the thalassaemia kit should be
	anned dad
	provided.
26	The system should have an on board QC Menu capable of storing the quality
.26	The system should have an on board QC Menu capable of storing the quality
	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation
Market Market	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values.
	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb
Market Market	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide.
Market Market	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms.
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result.
27 28 29	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result.
27	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic
27 28 29	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility.
27 28 29	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time.
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment.
27 28 29 30	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment.
27 28 29 30 31	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment.
27 28 29 30 31	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of
27 28 29 30 31 32	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption.
27 28 29 30 31	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption.
27 28 29 30 31 32 33	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption. The reagent containers should have a capacity of more than 1.5 liters so that the user does not need to change buffers regularly.
27 28 29 30 31 32	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2. Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption. The reagent containers should have a capacity of more than 1.5 liters so that the user does not need to change buffers regularly. The system must have a dual Piston Pump to give a continuous and a precise
27 28 29 30 31 32 33 34 35	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption. The reagent containers should have a capacity of more than 1.5 liters so that the user does not need to change buffers regularly. The system must have a dual Piston Pump to give a continuous and a precise buffer gradient.
27 28 29 30 31 32 33	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption. The reagent containers should have a capacity of more than 1.5 liters so that the user does not need to change buffers regularly. The system must have a dual Piston Pump to give a continuous and a precise buffer gradient.
27 28 29 30 31 32 33 34 35	The system should have an on board QC Menu capable of storing the quality control data and printing the standard deviation and Coefficient of Variation values. The company should provide normal and abnormal third party controls for Hb A2, Hb F and Hb S and provide External Quality Assurance Scheme (EQAS) to help compare results with similar users worldwide. The system should give information on the subtype and quantity of hemoglobin detected. Also the software should enable result storage of minimum 10000 chromatograms. The result from the machine should be presented in a symmetrical order (vertical) with proper description of date, time of injection, sample ID, age, sex total area count, different fractions of haemoglobins along with their quantity with flagging for out of range values and the chromatogram with each peak marked with their respective retention time for easy viewing of the result. It should have a built in vacuum-based degassing system, automatic equilibration and wash procedures and have built in column thermostat for reproducibility. The system must be capable of holding 10 samples at a time and should have the facility of expanding it to 50 samples along with a rack laoder so that it can be used for at least 50 samples at a time. The system should have a polyethylene waste tank which has a sensor to detect a 95% full tank and gives an alarm when sensor is tripped, as well as built in alarms for calibration and control failures for equipment. The waste tank should be sufficiently big (atleast 5 liters) so that it redues user interference with the machine and help in smooth running of large volume of samples without interruption. The reagent containers should have a capacity of more than 1.5 liters so that the user does not need to change buffers regularly. The system must have a dual Piston Pump to give a continuous and a precise buffer gradient.

	specialist and thalassemia expert for the technical and chromatogram interpretation related issues.
38	Standardization: Traceable to the Diabetes Control and Complications Trial
30	(DCCT) reference methods and IFCC. Certified via the National
	Glycohemoglobin Standardization Program (NGSP).
39	The manufacturer should be ISO 13485 certified.
40	The HPLC system should have flexibility to use various samples tube sizes of
70	13x75mm, 13x100 mm micro capillary tubes (Micro cuvette startstedt), Micro
	capillary tubes brands (other than Micro cuvette startstedt) defined as sample
	vials 1.5ml sample vials.
41	The HPLC system shoud have touch screen LCD display operation with
	resistive colour touch screen, simple to use, user friendly and with menu driven
	facility. The software should be able to show not just the values and run process
	but also full chromatogram.
42	The HPLC system should have built in thermal printer along with external
	printer connected.
43	The instrument provider should also provide training to the pathologist for
	working, quality control, testing, evaluation and interpretation.
38	Standardization: Traceable to the Diabetes Control and Complications Trial
	(DCCT) reference methods and IFCC. Certified via the National
	Glycohemoglobin Standardization Program (NGSP).
39	The manufacturer should be ISO 13485 certified.
40	The HPLC system should have flexibility to use various samples tube sizes of
	13x75mm, 13x100 mm micro capillary tubes (Micro cuvette startstedt), Micro
	capillary tubes brands (other than Micro cuvette startstedt) defined as sample
	vials 1.5ml sample vials.
41	The HPLC system shoud have touch screen LCD display operation with
	resistive colour touch screen, simple to use, user friendly and with menu driven
	facility. The software should be able to show not just the values and run process
	but also full chromatogram.
42	The HPLC system should have built in thermal printer along with external
	printer connected.
43	The instrument provider should also provide training to the pathologist for
	working, quality control, testing, evaluation and interpretation.

FORMAT OF TECHNICAL COMPLIANCE OF SPECIFICATIONS

Signature & seal of bidder

• If the bidder fails to enclose the compliance Statement, his bid is likely to be rejected.

Deviation Statement Form

The following are the particulars of the deviations from the requirements of the Tender Document terms and conditions:

Sl.	Clause in the terms and conditions	Deviation	Justification for the deviation
140.			

Annexure-D (Part of price bid)

Cost of the Equipment, Parts and Accessories and other details offered as per the suggested mandatory components

(Must be enclosed in the Price Bid)

Sl. No.	Item	Unit Price	Amount (Currency)
	Specifications for Equipments		

Annexure-D(1)

(To be Attached along-with Price Bid only)

Installation, Warranty & Supply of Parts/Spares

S1.	Particulars	Description/Annexure
No.		
1	Free Warranty Details	
2	Number of Preventive Maintenance	
	Visits per year during FREE Warranty	
3	Discount to be offered by the company	
	on the spares/parts/consumables/	
	attachments etc. for the period of	
	minimum 05 years after supply and	
	installation of the equipment	
4	Whether the charges for Installation,	
	Commissioning and testing at NIRTH,	
	Jabalpur are included in the price list.	

Authorized signature with stamp

(To be Attached along-with Price Bid only)

AMC/CAMC Details

Year-wise AMC/CAMC charges	Year	AMC charges	CAMC charges
For 05 years. (after free warranty)	1 st year		
	2 nd year		
	3 rd year		
	4 th year		
	5 th year		

Authorized signature with stamp

After Sales Service Support Details

SI. No.	Local Address, Telephone numbers, Fax numbers and email address of the agent who will provide after sales support	Number of engineers/service personnel who are stationed at the local address	Nature of training imparted to the service personnel especially at the principal's company

The bidder should clearly indicate in which cases it would be essential for the equipment to be sent to foreign principal for servicing or in which cases it will be essential for service personnel to come from foreign principal.

WARRANTY CERTIFICATE

We Warrant that everything to be supplied by us hereunder shall be brand new, free from all defects and faults in materials, workmanship and manufacture and shall be of the highest grade and quality and consistent with the established and generally accepted standards for material of the type ordered shall be in full conformity with the specification, drawing, or samples, and operate properly. We shall be fully responsible for its efficient operation. This Warranty shall survive inspection of any payment or and acceptance of the goods but shall expire after (expect in respect of complaints of which the supply has been notified prior to such date) 36 months Comprehensive warranty and 24 months non-comprehensive warranty, if equipment cost is more than 10.00 lakhs otherwise 3 years comprehensive warranty, after their successful installation, commissioning and acceptance by ICMR-NIRTH, JABALPUR.

The obligations under the Warranty expressed above shall include all costs relating to labor, spares, maintenance (preventive and unscheduled) and transport charges from site to the manufacturers work and back and free repair/adjustment or replacement at site or any parts of the equipment which under normal and proper use and maintenance proves defective in design, material or workmanship or fails to conform to the specifications previously given by the **NIRTH**, JABALPUR to the supplier.

Signature & seal of bidder

Note: It should be provided in the letter head of the firm/ company.

BID SECURITY/EARNEST MONEY DEPOSIT FORM

Whereas				
KNOWE ALL PEOPLE by these presents that WE				
Sealed with the Common Seal of the said Bank this day of 20 THE CONDITION of this obligation are:				
1.If the Bidder withdraws its Bid during the period of bid validity specified by the Bidder on the Bid Form; or 2. If the Bidder, having been notified of the acceptance of its bid by the Purchaser during the period of bid validity:				
 a) fails or refuses to execute the Contract Form if required; or b) fails or refuses to furnish the performance security, in accordance with the instructions of Bidders. 				
We undertake to pay the Purchaser up-to the above amount upon receipt of its first written demand, without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due to it, owing to the occurrence of one or both of the two conditions, specifying the occurred condition or conditions.				
This guarantee shall remain in force up-to and including forty five (45) days after the period of the bid validity, and any demand in respect thereof should reach the Bank not later than the above date.				
(Signature of the Bank)				

1 Name of Bidder

Check List to be furnished by the Bidder

- 1) When Quotation is from Indian agent / Dealer whether letter of Authorization of Foreign principal is enclosed in the techno-commercial bid Yes / No.
- 2) For Two Bid System Whether Techno Commercial and Price Bids are kept in separate envelops duly indicating Techno Commercial Bid and Price Bid respectively- Yes / No.
- 3) Whether Tender Number, due date and opening date have been written in all the envelops Yes / No.
- 4) In case of Tenderers who downloaded the documents from the website Whether the tender document fee (Demand draft in favor of Director National Institute for Research in Tribal Health, Jabalpur is enclosed along with Techno commercial bid Yes / No.
- 5) EMD Whether the required EMD is enclosed along-with the Techno commercial bid Yes / No.
- 6) In case the firm is exempted from payment of EMD (like SSIS, registered with NSIC, MSEs, firms registered under DGS&D) Whether Valid Permanent Enlistment Certificate of NSIC/Valid DGS&D Rate Contract copy is enclosed in the Techno Commercial bid-Yes / No.
- 7) If You are an Indian Agent / Dealer / Distributor for a foreign principal have you enclosed the Letter of Authorization / Agency Certificate in the Techno Commercial Bid Yes / No.
- 8) Whether Technical compliance Statement in the format given in Annexure A is enclosed in the Techno Commercial Bid Yes / No.
- 9) Whether the Deviation Statement in the format given in Annexure B is enclosed in the Techno Commercial Bid Yes / No.
- 10) Whether Details of after sales support is given as per format in Annexure-D and is enclosed in the Techno Commercial Bid Yes / No.
- 11) Whether the list of users is given as per Annexure K and is enclosed in the Techno Commercial Bid Yes / No.
- 12) Whether schedule of AMC charges for a period of 05 years has been attached along-with the price list in Annexure-D(1) Yes / No.
- 13) Whether the warranty certificate is enclosed as per Annexure F and is enclosed in the Techno Commercial Bid Yes / No.
- 14) Whether the Pre-installation requirements is given in the Techno Commercial Bid Yes / No.
- 15) Whether the Price bid is as per the format given in Annexure C Yes / No.
- 16) Whether the Price Bid has been signed in every page Yes / No.
- 17) Whether audited balance sheet of last three years are enclosed with all relevant documents Yes / No.

Annexure I

List of Users for a period of 3 years preceding this tender (if possible similar equipment)

Sl.	Order for	Purchase	Description,	Price charged	Date of	Reason of any,
No.	similar equipment placed by (full address of purchaser with contact person,	Order Number and Date	Model and quantity ordered	Important, (Rates to be blacked out in the technical bid The Price bid of bidder must contain the price charged)	completion of delivery as per contract and Date of installation/ commissioning as per contract	for delay in delivery, installation and commissioning,
	telephone number)					

Performance Bank Guarantee

Bank Guarantee No.:
Amount of Guarantee: Rs Valid Up to :
Guarantee Cover from:to
The Director, ICMR-National Institute for Research in Tribal Health, Ministry of Health, Govt. of India Nagpur Road, P.O. Garha, Jabalpur – 482003
This deed of guarantee executed by (Name of Bank)
ICMR-National Institute for Research in Tribal Health – Jabalpur-482003. (hereinafter referred to as "the Beneficiary") for an amount not exceeding Rs. (Rupees
Contractors").
This Guarantee is issued subject to the condition that the liability of the Bank under this Guarantee is limited to a maximum of Rs
Where as you have entered in to a contract reference No. P.O. No
the supply of
said contract" and whereas (Name of Firm)
ICMR-National Institute for Research in Tribal Health – Jabalpur – 482003. for the warranty of the equipment supplied.
We, (Name of Bank)
that in the event ICMR- National Institute for Research in Tribal Health, Jabalpur – 482003 declares to us through you that (Name of Firm) has not fulfilled the warranty/other obligations according to the contractual warranty obligations under the said contract to pay you on demand without any reference to M/s an amount of Rs. (Rupees
or proceedings pending in any competent Indian Court or before any arbitration tribunal, your written demand shall be conclusive evidence to us that such payment is payable under the terms of the said contract and shall be binding in all respect on us.
We shall not be discharged or released from the aforesaid undertaking and guarantee by any arrangements, variations made between you and (Name of Firm)

by any forbearance whether as to payment time performance or otherwise.				
We further agree and undertake not to revoke this guarantee before the same is discharged in writing by you.				
This guarantee shall remain valid until				
 Notwithstanding anything contained herein. Our liability under this guarantee is restricted to Rs				
Dated: Place:				
(Signature of authorized Officer of the Bank)				
Witness (Name, Address & Signature)				
1.				
2.				

MANDATE FORM FOR RTGS/NEFT

Annexure-K

To, The Director, ICMR-NIRTH, Nagpur Road, Garha PO, Jabalpur-482 003 (M.P.)

	III 1 N (D C' N)	
1	Vendor Name (Beneficiary Name)	
2	PAN Number	
3	GSTN	
4	Mobile No.	
5	Phone No.	
6	Email	
7	Complete Address	
	City	
	Country	
	State	
	District	
	Pin Code Number	
8	Bank Name	
9	Account Number	
10	IFSC Code	

NOTE: Please attach a photocopy of cancelled cheque for purpose of verification of the concerned Bank account.

I hereby declare that the particulars given above are correct and complete. If the transaction is delayed or not affected at all for reasons of incomplete and incorrect information, I would not be hold the user Institute responsible.

Signature/Seal of Dealer/Firm

Date:

General Guidelines for furnishing breakup

Total cost of Basic Equipment / System (In case of imports without incidence or Customs Duty)

- Ex Works Prices
- Cost of Handling, Inland Transportation, Insurance charges etc. /FOB Charges (Nearest International Airport)
- Estimated Cost of Air / Ocean Freight, Insurance, transportation etc. up-to New Delhi/Mumbai airport/seaport/CIF Price New Delhi/Mumbai Shipment by Air/Sea to be clearly mentioned)
- Agency Commission payable to Indian Agent as percentage / annuity NIRTH, Jabalpur will
 pay the Indian Agent in Indian Rupee only after satisfactory installation and commissioning
 of the Equipment Agency commission is calculated from the Ex works Price).

Or

Total cost of the Basic Equipment / System (In case of Indigenous offers – Tenderers should quote without excise duty as this Institution is exempted from payment of excise duty under notification 10/97)

- Ex Works Prices
- Cost of packing, forwarding, freight, insurance charges etc. up-to Jabalpur.
- Taxes if any,
 - Cost of Each Spare/Part/Accessory
 - Banking Charges (in case of Imports)
 - In case of Letter of Credit:
- Banking charges inside India and Outside India (To whose account to be clearly indicated for Establishing LC)
- Confirming Bank's Address) if any, to be clearly indicated In case of Telegraphic Transfer /Swift Transfer:
- Address of Banker with Account Number and Swift Code
 - Training costs (if any):
 - In India/Abroad. (Please indicate place, duration and number of persons included for training)

Additional Information

Sl. No.	Item	Terms of bidder
1.	Payment Terms	
2.	Delivery Period	
3.	Installation and Commissioning	
4.	Country of Origin	
5.	Port of shipment (By Air/Sea to be clearly indicated	
6.	Net weight of the equipment and gross weight after packing	
7.	Net Dimensions of the equipment and gross dimensions after packing (with number of packages)	
8.	Company Name and Address to whom the order has to be placed	
9.	AMC Charges per year after warranty period	